

VisiPro DO Ex 120 H2



Product Identification

Ref	243667-112
SN	4153
Lot	111093215
Heat Number	Shaft 549384
	Cap 812028
Firmware	ODOUH008

Factory Test Conditions

Tests were performed according to	606130
Temperature	24.5 °C

Factory Test Results

	Test Specification	Result
Phase Φ in air (20.95 %-vol.; 100 %-sat)	37.0 ° \pm 3.0 °	35.8 °
Phase Φ at zero point (0 %-vol.; 0 %-sat)	71.0 ° \pm 2.0 °	72.1 °
Quality Indicator	100 %	passed
Functionality Test (HART, 4-20 mA)		passed
Quality Test successfully passed	2019-10-07	